



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/637,529	08/11/2000	Robert C. Beck	1480	8331

7590 04/07/2006

ROBERT C. BECK
BECK & TYSVER
2900 THOMAS AVE S #100
MINNEAPOLIS, MN 55416-4463

EXAMINER

DESANTO, MATTHEW F

ART UNIT	PAPER NUMBER
----------	--------------

3763

DATE MAILED: 04/07/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No. 09/637,529	Applicant(s) BECK, ROBERT C.	
	Examiner Matthew F. DeSanto	Art Unit 3763	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 17 November 2005.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 19 and 21-26 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 19 and 21-26 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:


A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

2. Claims and 19, 21, 22 are rejected under 35 U.S.C. 102(e) as being anticipated by Drasler et al. (USPN 5,496,267).

Drasler et al. discloses a fluid supply catheter, a lumen, a distal aperture, and a sheath, wherein the ablation catheter is located within the sheath and adapted for motion with respect to the sheath and wherein said aperture defining a first aperture defining a first aperture direction for the emerging flow that lies between approximate zero degrees and ninety degrees. (Figure 14 and the respected text in the reference)



3. Claim 26 are rejected under 35 U.S.C. 102(e) as being anticipated by Ruggio (USPN 5,476,450).

Ruggio discloses a fluid supply catheter, a lumen, a distal aperture (424), a control body (430) and a sheath, wherein the ablation catheter is located within the sheath and adapted for motion with respect to the sheath and wherein said aperture defining a first aperture defining a first aperture direction for the emerging flow that lies between approximate zero degrees and ninety degrees. (Figures 12, 13, 14, 15 and entire reference)

Art Unit: 3763

Claim Rejections - 35 USC § 103

4. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

5. Claims 23-25 are rejected under 35 U.S.C. 103(a) as being unpatentable over Kensey (USPN 4,631,052) and further in view of Ruggio (USPN 5,476,450).

Kensey discloses a catheter that injects fluid through an aperture (50) at a first rate, and wherein the fluid aperture is located proximate a control body (58a) removing fluid and entrained flow through said fluid port (56), but fails to disclose the use of a delivery sheath.

Ruggio discloses the use of a delivery sheath and a therapeutic catheter for administering treatment to the patient (see figures 12-15).

Therefore, at the time of the invention it would have been obvious for one of ordinary skill in the art to combine the teachings of Kensey with Ruggio because Ruggio teaches the use of a large tube to be used since this allows for easier navigation than using several smaller tubes, since the larger tube allows several smaller tubes to be passed within the larger tube, and thus be able to be navigated to the treatment area (column 10, 11).

Double Patenting

6. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent

Art Unit: 3763

and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

7. Claims 19, 21-26 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-16 of U.S. Patent No. 6,129,698. Although the conflicting claims are not identical, they are not patentably distinct from each other because the USPN 6129698 teaches a catheter body, apertures, and a control body.

Response to Arguments

1. Applicant's arguments filed 11/17/05 have been fully considered and are persuasive with regards to Neracher (USPN: 5135482), Drasler and Ruggio. Therefore, new grounds of rejection have been made with new interpretation of the prior art.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Matthew F. DeSanto whose telephone number is 571-272-4957. The examiner can normally be reached on Monday-Friday 9:30-6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick LUCCHESI can be reached on (571) 272-4977. The fax phone

Art Unit: 3763

number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Matthew DeSanto
Art Unit 3763
April 3, 2006



NICHOLAS D. LUCCHESI
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700